



Phosphate LR T

M319

0.05 - 4 mg/L PO<sub>4</sub>PO<sub>4</sub>

Phosphomolybdenum Blue

## Instrument specific information

The test can be performed on the following devices. In addition, the required cuvette and the absorption range of the photometer are indicated.

Instrument Type	Cuvette	$\lambda$	Measuring Range
PM 600, PM 620, PM 630	ø 24 mm	610 nm	0.05 - 4 mg/L PO <sub>4</sub>

## Material

Required material (partly optional):

Reagents	Packaging Unit	Part Number
Phosphate No. 1 LR	Tablet / 100	513040BT
Phosphate No. 2 LR	Tablet / 100	513050BT
Phosphate No. 2 LR	Tablet / 250	513051BT
Set Phosphate No. 1 LR/No. 2 LR 100 Pc. #	100 each	517651BT

## Application List

- Waste Water Treatment
- Boiler Water
- Drinking Water Treatment
- Raw Water Treatment
- Pool Water Control



## Preparation

1. Strongly buffered samples or samples with extreme pH values should be adjusted to between pH 6 and pH 7 before the analysis (use 1 mol/l Sulphuric acid or 1 mol/l Sodium hydroxide).
2. Ortho-Phosphate ions react with the reagent to form an intense blue colour. Phosphate, which is found in organic and condensed, inorganic (meta-, pyro- and polyphosphate) forms, must therefore be converted into ortho-phosphate ions prior to analysis. The pretreatment of the sample with acid and heat creates the conditions for the hydrolysis of the condensed, inorganic forms. Organically bound phosphate can be converted into ortho-phosphate ions by heating with acid and Persulphate.

The amount of organically bound phosphate can be calculated:  
 $\text{mg/L organic Phosphate} = \text{mg/L Phosphate, total} - \text{mg/L Phosphate, can be hydrolysed in acid.}$

## Notes

1. Only ortho-phosphate ions react.
2. The tablets must be added in the correct sequence.



## Determination of Phosphate, ortho LR with Tablet

Select the method on the device.


For this method, a ZERO measurement does not have to be carried out every time on the following devices: XD 7000, XD 7500



Fill 24 mm vial with **10 mL sample**.



Close vial(s).



Place **sample vial** in the sample chamber. Pay attention to the positioning.



Press the **ZERO** button.



Remove the vial from the sample chamber.


For devices that require **no ZERO measurement**, start here.




Add **PHOSPHATE No. 1 LR tablet**.




Crush tablet(s) by rotating slightly.




Add **PHOSPHATE No. 2 LR tablet**.



Crush tablet(s) by rotating slightly.




Close vial(s).




Dissolve tablet(s) by inverting.



Place **sample vial** in the sample chamber. Pay attention to the positioning.



Press the **TEST** (XD: **START**) button.



Wait for **10 minute(s) reaction time**.

Once the reaction period is finished, the measurement takes place automatically. The result in mg/L ortho-Phosphate appears on the display.



## Analyses

The following table identifies the output values can be converted into other citation forms.

Unit	Cite form	Scale Factor
mg/l	P	0.3261
mg/l	PO <sub>4</sub> <sup>3-</sup>	1
mg/l	P <sub>2</sub> O <sub>5</sub>	0.7473

## Chemical Method

Phosphomolybdenum Blue

## Appendix

### Interferences

Interference	from / [mg/L]
Al	200
AsO <sub>4</sub> <sup>3-</sup>	in allen Mengen
Cr	100
Cu	10
Fe	100
Ni	300
H <sub>2</sub> S	in allen Mengen
SiO <sub>2</sub>	50
S <sup>2-</sup>	in allen Mengen
Zn	80
V(V)	große Mengen
W(VI)	große Mengen

### According to

DIN ISO 15923-1 D49  
Standard Method 4500-P E  
US EPA 365.2

\* including stirring rod, 10 cm