

Chlorine Dioxide LR

56I700220

0.16 - 12 mg/L ClO₂

Material

Reagents	Packaging Unit	Part Number
Chlorine Dioxide Buffer CDO1	65 mL	56L033965
Chlorine Dioxide Titrant CDO2A	65 mL	56L150165

The following accessories are required.

Accessories	Packaging Unit	Part Number
Syringe, plastic, 20 mL	1 Pieces	56A006501
Titration jar with cap, plastic, 250 mL	1 Pieces	56A010501

Application List

- Cooling Water


Remarks

1. Colours may vary depending on sample and test conditions.
2. Acidic samples must be neutralised prior to testing.


Sampling

Select the sample volume from the table according to the expected measuring range and read off the factor to calculate the result.

Expected Range	Titrant used	Sample Size	Factor
0.16-0.48 mg/L	Chlorine Dioxide Titrant CDO2A	250 mL	0.02
0.2-0.6 mg/L	Chlorine Dioxide Titrant CDO2A	200 mL	0.025
0.4-1.2 mg/L	Chlorine Dioxide Titrant CDO2A	100 mL	0.05
0.8-2.4 mg/L	Chlorine Dioxide Titrant CDO2A	50 mL	0.1
1.0-3.0 mg/l	Chlorine Dioxide Titrant CDO2A	40 mL	0.125
2.0-6.0 mg/L	Chlorine Dioxide Titrant CDO2A	20 mL	0.25
4.0-12 mg/L	Chlorine Dioxide Titrant CDO2A	10 mL	0.5




Attention! Select the appropriate sample volume according to the instructions in the chapter Sampling.



Add **25 drops Chlorine Dioxide Buffer CDO1**.




Swirl to mix.




Attention! Record the number of drops that will be added.

Note: Make sure to swirl the jar after adding each drop!



Add **Chlorine Dioxide Titrant CDO2A** drop by drop to the sample until colouration turns from **colourless/pale yellow to blue/green**.



Swirl to mix. Allow 5 seconds between each addition.

Calculate test result: Chlorine Dioxide (as ClO_2) mg/L = Number of drops x factor (see table)